

Extending the Range of Rate Constants Available from BIACORE: Interpreting Mass Transport-Influenced Binding Data

David G. Myszka,* Xiaoyi He,# Micah Dembo,§ Thomas A. Morton,[¶] and Byron Goldstein[#]

*Huntsman Cancer Institute, University of Utah, Salt Lake City, Utah 84112-5330 USA; #Theoretical Biology and Biophysics Group, Theoretical Division, Los Alamos National Laboratory, Los Alamos, New Mexico 87545 USA; §Department of Bioengineering, Boston University, Boston, Massachusetts 02215 USA; and [¶]Division of Biochemistry and Molecular Biology, John Curtin School of Medical Research, Australian National University, Canberra, ACT 0200, Australia

ABSTRACT Surface-based binding assays are often influenced by the transport of analyte to the sensor surface. Using simulated data sets, we test a simple two-compartment model to see if its description of transport and binding is sufficient to accurately analyze BIACORE data. First we present a computer model that can generate realistic BIACORE data. This model calculates the laminar flow of analyte within the flow cell, its diffusion both perpendicular and parallel to the sensor surface, and the reversible chemical reaction between analyte and immobilized reactant. We use this computer model to generate binding data under a variety of conditions. An analysis of these data sets with the two-compartment model demonstrates that good estimates of the intrinsic reaction rate constants are recovered even when mass transport influences the binding reaction. We also discuss the conditions under which the two-compartment model can be used to determine the diffusion coefficient of the analyte. Our results illustrate that this model can significantly extend the range of association rate constants that can be accurately determined from BIACORE.

INTRODUCTION

Until recently, obtaining reliable equilibrium and rate constants for interacting biomolecules was often difficult and time consuming. Optical biosensors now offer a rapid way to determine equilibrium and rate constants without the need to label the interacting biomolecules (Garland, 1996; Silin and Plant, 1997). Because of these features these instruments, such as BIACORE (Biacore, Uppsala, Sweden), have gained wide use and have been employed in the study of the interactions of a variety of biomolecules, including proteins, nucleic acids, lipids, and carbohydrates (reviewed in Szabo et al., 1995; Raghavan and Bjorkman, 1995; van der Merwe and Barclay, 1996; Myszka, 1997; Schuck, 1997a; Fivash et al., 1998).

In BIACORE instruments, one of the reactants is immobilized on a sensor chip. We will call it the receptor, in analogy to a receptor on a cell surface, although in the literature it is often referred to as the immobilized ligand. The other reactant, called the analyte, flows past the chip. The immobilization of the receptor is usually accomplished by coupling it to a thin dextran layer that extends ~100 nm out from the sensor surface, 0.2% of the height of the flow chamber. Detection of binding is based on the optical phenomenon of surface plasmon resonance (SPR). The SPR response is used to detect changes in the index of refraction caused by mass changes at the sensor surface. These changes are brought about by the binding of analyte to

receptor (Jönsson et al., 1991; Malmqvist, 1993; Garland, 1996). Thus continuous monitoring of the SPR signal allows the kinetics of binding to be followed in real time. After binding, buffer alone may be introduced to monitor the dissociation kinetics.

As the preceding description suggests, the SPR signal in a typical experiment is not simply a report of the progress of the chemical processes of association and dissociation at the surface. Rather, the signal is the result of a combination of these chemical processes and the transport processes of diffusion and flow. Thus obtaining estimates of the intrinsic forward and reverse rate constants of the analyte-receptor reaction requires some model that will, when needed, account for transport. The model most often used assumes that after a brief transient, during which analyte is transported to the sensor surface, no correction is needed, because the free analyte concentration remains uniform in space and constant in time, kept so by the continuous influx of new analyte. We will refer to this model as the rapid mixing model because, after the brief transient, the binding at the sensor surface is the same as for a well-mixed system with constant analyte concentration.

Systematic errors in the estimates of the rate constants can arise if the rapid mixing model is not a reasonable description of the binding kinetics (Chaiken et al., 1992; Malmberg et al., 1992; Ito and Kurosawa, 1993; Morton et al., 1994; Wohlhueter et al., 1994). A number of authors have discussed this problem and presented conditions under which the assumption is expected to break down (Glaser, 1993; Karlsson et al. 1994; Schuck, 1996; Schuck and Minton, 1996; Yarmush et al., 1996; Christensen, 1997). There are two possible sources for the breakdown: 1) the transport (diffusion and flow) of analyte in solution to the dextran layer and 2) the transport (primarily diffusion) of

Received for publication 30 December 1997 and in final form 29 April 1998.

Address reprint requests to Dr. Byron Goldstein, Theoretical Biology and Biophysics Group, Theoretical Division, T-10, MS K710, Los Alamos National Laboratory, Los Alamos, NM 87545. Tel.: 505-667-6538; Fax: 505-665-3493; E-mail: bxg@lanl.gov.

© 1998 by the Biophysical Society

0006-3495/98/08/583/12 \$2.00

analyte within the dextran layer. If one or both of these influence the kinetics of binding, then the free analyte concentration cannot be taken to be constant and uniform.

In this paper we consider only experiments in which the properties of the dextran layer do not influence the binding kinetics, i.e., where the binding can be treated as if the receptors are coupled directly to the sensor surface. BIA-CORE binding studies have been performed both with receptors coupled to dextran layers and with receptors coupled directly to the sensor surfaces (Karlsson and Fält, 1997; Parsons and Stockley, 1997). These experiments demonstrate that under appropriate experimental conditions (i.e., low surface capacity), the dextran layer has no significant effect on the binding kinetics.

To improve the analysis of bulk transport effects on the SPR signal, a simple model has been proposed that treats binding as a two-step process: transport of analyte to the sensor surface, followed by reaction of analyte with receptors on the surface (Myszka et al., 1997). This two-compartment model is attractive because it can be formulated in terms of a simple system of ordinary differential equations and because it is only slightly more complicated than the rapid mixing model. The two-compartment model has been used to describe several experimental systems studied with BIACORE instruments (Myszka et al., 1996, 1997; Morton and Myszka, 1998). This is encouraging, but uncertainty remains about how well the two-compartment model describes BIACORE binding kinetics. What is provided in this paper is a demonstration that the two-compartment model is sufficiently accurate to analyze BIACORE data. In particular, we show that the numerical parameter values obtained by fitting the two-compartment model to noisy SPR data are accurate estimates of the underlying physical parameter values.

The approach we use to demonstrate the validity of the two-compartment kinetic scheme for analyzing BIACORE binding data is as follows:

1. We present a computer model that can simulate BIACORE binding experiments involving monovalent analytes binding to monovalent receptors, i.e., that accounts for laminar flow and diffusion of the analyte, chemical reactions at the sensor surface, and geometry of the BIACORE. This computer model is similar to that developed by Glaser (1993), the difference being that we include diffusion in the flow direction.

2. We use the computer model to simulate BIACORE binding data, choosing parameter values that lead to transport-influenced binding kinetics as well as parameter values where such effects are negligible.

3. We fit the two-compartment binding model to the simulated data and show that we recover the original values of the rate constants (the values used to simulate the data).

4. We then add random noise at realistic levels to the simulations and demonstrate that good estimates of the input parameters are recovered when the two-compartment model is used to analyze the noisy simulated data.

OBTAINING SIMULATED DATA SETS

The mathematical model

In BIACORE instruments, analyte is transported by diffusion and flow to the sensor surface, where it reacts with immobilized receptors. The flow channel of the BIACORE has a rectangular cross section (length l , height h , and width w). We let t be the time, and we introduce Cartesian coordinates with the origin at the inflow port, the x axis parallel to the direction of flow, and the y axis normal to the receptor-coated surface (illustrated in Fig. 1). (In BIACORE instruments, the sensor surface is actually on top of the flow cell. For convenience we draw it on the bottom.) Because the instrument's flow chamber is 10 times wider than it is high, variations in concentrations across the width of the chamber can be neglected.

Along the flow chamber, laminar flow is fully developed essentially over its entire length (Brody et al., 1996). The velocity profile is therefore parabolic, equal to zero at the top ($y = h$) and bottom ($y = 0$) boundaries and rising to a maximum, v_c , in the center, i.e., the velocity $v(y)$ at a height y above the sensor surface is (Batchelor, 1967)

$$v(y) = 4v_c(y/h)(1 - (y/h)) \quad (1)$$

(Because Eq. 1 is sometimes written in terms of the average velocity, \bar{v} , we note that $\bar{v} = 2v_c/3$.)

The dependent variables of the model are the bulk concentration of free analyte, $C(t, x, y)$, and the surface density of bound analyte, $B(t, x)$ (with units of mass/volume and mass/area, respectively). The net analyte flux due to convection and diffusion is

$$\mathbf{J}(t, x, y) = \mathbf{i}v(y)C(t, x, y) - D(\mathbf{i}\partial C(t, x, y)/\partial x + \mathbf{j}\partial C(t, x, y)/\partial y) \quad (2)$$

where \mathbf{J} is the flux vector, D is the analyte diffusion coefficient, and \mathbf{i} and \mathbf{j} are the unit vectors in the x and y directions. The component of \mathbf{J} normal to any surface, integrated over that surface, equals the flux through the surface, e.g., the number of particles per second passing through the surface. From conservation of mass (the continuity equation), the following partial differential equation




FIGURE 1 Schematic of a BIACORE flow chamber, where the variation along the width of the chamber has been neglected and the sensor surface has been placed on the bottom rather than the top. For a standard instrument, $h = 5 \times 10^{-3}$ cm, $l = 0.24$ cm, and the width equals 5×10^{-2} cm. The figure is not to scale. The true length is 48 times longer than the height.

(PDE) is obtained:

$$\partial C/\partial t = D(\partial^2 C/\partial x^2 + \partial^2 C/\partial y^2) - 4v_c(y/h)(1 - (y/h))\partial C/\partial x \quad (3)$$

The boundary conditions for Eq. 3 are as follows. Because the top boundary is impenetrable and unreactive, the mass flux of analyte must vanish at this boundary:

$$\partial C(t, x, y)/\partial y = 0 \quad \text{at } y = h \quad (4a)$$

In contrast, the analyte mass flux at the sensor surface should equal the time rate of change in the amount bound at the surface, i.e.,

$$D\partial C(t, x, y)/\partial y = \partial B(t, x)/\partial t = k_a C(t, x, 0)R(t, x) - k_d B(t, x) \quad \text{at } y = 0 \quad (4b)$$

where $C(t, x, 0)$ is the free analyte concentration at position x just above the sensor surface. $R(t, x) = R_T - B(t, x)$ is the free surface receptor concentration at x . R_T is the total receptor concentration and is constant with respect to both position and time. The rate constants of the reaction are k_a and k_d . Note that these are the intrinsic microscopic forward and reverse rate constants for binding of the analyte to the receptor.

At the entry port to the flow cell ($x = 0$), the concentration is constant and equal to the injection concentration, C_T , i.e.,

$$C(t, 0, y) = C_T \quad \text{at } x = 0 \quad (4c)$$

At the end of the sensor, where analyte exits ($x = l$), we adopt a simple continuation condition that assumes the exit of analyte is due entirely to flow:

$$\partial C(t, x, y)/\partial x = 0 \quad \text{at } x = l \quad (4d)$$

Hence they leave the computational domain and have a negligible effect on the dynamic processes happening inside the instrument.

It is useful to write Eq. 3 and the boundary conditions, Eqs. 4a–4d, in nondimensional form. To do this we introduce the following nondimensional time and lengths, concentrations, and rate constants:

$$\tau \equiv Dt/h^2 \quad x \equiv x/l \quad y \equiv y/h \quad (5a)$$

$$c \equiv C/C_T \quad b \equiv B/R_T \quad r \equiv R/R_T \quad (5b)$$

$$\kappa_a \equiv k_a C_T h^2/D \quad \kappa_d \equiv k_d h^2/D \quad (5c)$$

and nondimensional parameters:

$$\epsilon = h/l \quad p = 4v_c h/D \quad (5d)$$

Equations 3 and 4a–4c become

$$\partial c/\partial \tau = \epsilon^2 \partial^2 c/\partial x^2 + \partial^2 c/\partial y^2 - p\epsilon y(1 - y)\partial c/\partial x \quad (6)$$

$$\partial c(\tau, x, y)/\partial y = 0 \quad \text{at } y = 1 \quad (7a)$$

$$(hC_T/R_T)\partial c(\tau, x, y)/\partial y = \partial b(\tau, x)/\partial \tau = \kappa_a c(\tau, x, 0)(1 - b(\tau, x)) - \kappa_d b(\tau, x) \quad \text{at } y = 0 \quad (7b)$$

$$c(\tau, 0, y) = 1 \quad \text{at } x = 0 \quad (7c)$$

$$\partial c(\tau, x, y)/\partial x = 0 \quad \text{at } x = 1 \quad (7d)$$

If diffusion in the x direction is ignored (i.e., the ϵ^2 term in Eq. 6 is set equal to zero), the model reduces to what has now become a standard description for the BIACORE flow cell (Lok et al., 1983; Glaser, 1993; Christensen, 1997). Yarmush et al. (1996) have also used this approximation, but in the context of a nonnegligible dextran layer. For a standard flow cell, $\epsilon = 0.021$. This suggests that such an approximation is justified, although for $y < \epsilon$, diffusion in the x direction will dominate flow. We have not made this approximation, because we are using the model as an absolute standard for generating simulation data, and therefore it is desirable to include as much detail as possible. Because we are solving Eq. 6 numerically, there is no necessity to do otherwise.

The numerical method

We have solved Eq. 6 by using a general, conservative, and well-tested hydrodynamics transport code based on the method of finite elements (Dembo, 1994a,b). Shown in Fig. 2 is a small computational grid (8×8) with 49 node points and 32 surface elements. In the x direction the length is divided uniformly into eight parts. In the y direction the height first is divided into four equal parts. The top and bottom sections are then divided in half, so there are six sections in all. Finally, the new top and bottom sections are divided in half, forming the eight sections shown in Fig. 2. The height of the bottom and top sections is $h/16$. Refinements of this grid are generated by subdividing each of the quadrilaterals of the course mesh into four equal parts. For example, the top and bottom elements of a 64×64 grid are $h/128$ in height and $l/64$ in length. In the BIACORE flow cell, except at very short times after the initiation of binding and dissociation, the concentration of analyte is uniform far from the sensor surface. Thus there is no reason for generating a fine grid near the top surface. This was done only because the code we adapted required a grid that was




FIGURE 2 An 8×8 computational grid consisting of 49 node points and 32 boundary points.

symmetrical about $h/2$. We have since developed a code for which this is not required (unpublished result).

In trial simulations we compared the time course for the average concentration of bound analyte, calculated using a 32×32 and a 64×64 grid. The average concentration of the bound ligand was calculated by summing the bound ligand concentration over each bottom surface element and dividing the sum by the number of bottom surface elements, $n + 1$ for an $n \times n$ grid. There was no significant difference between the values obtained using the two grids. Nevertheless, in the Results section, where we test the two-compartment model's ability to recover accurate parameter values, all simulations were done on a 64×64 grid. For this grid the first set of grid points above the sensor surface is at $y = h/128 \approx 400$ nm, or about four times the thickness of the dextran layer for a standard CM5 sensor surface (Stenberg et al., 1991).

Simulations

To illustrate the information that can be obtained from solving Eq. 6 numerically, subject to the boundary conditions in Eqs. 7a–7d, we simulate a BIACORE binding study using parameter values previously determined for interleukin-2 (IL-2), flowing past, and interacting with, its immobilized low-affinity receptor, IL-2R α (Myszka et al., 1996). In the simulation we took the association rate constant $k_a = 8 \times 10^6 \text{ M}^{-1} \text{ s}^{-1}$ and the dissociation rate constant $k_d =$




FIGURE 3 Predicted time course for the average concentration of bound analyte (—) and average free analyte concentration at the sensor surface (---). In the simulation, 25 nM analyte was injected from 0 to 50 s ($C_T = 25$ nM), then the injected concentration was set to zero ($C_T = 0$) and dissociation was followed for 50 s. $v_c = 5$ cm/s, corresponding to a flow rate of $50 \mu\text{l}/\text{min}$, and the analyte diffusion coefficient $D = 1 \times 10^{-6} \text{ cm}^2/\text{s}$. The immobilized receptor concentration $R_T = 7.5 \times 10^{11}$ receptors/ $\text{cm}^2 = 1.25$ nM cm. (Because $1 \text{ RU} = 10^{-10} \text{ g}/\text{cm}^2$, when all sites are filled with IL-2, MW = 14,000, there is an increase of 175 RU over baseline.) A standard flow cell was assumed, with $h = 5 \times 10^{-3} \text{ cm}$ and $l = 0.24 \text{ cm}$. $k_a = 8 \times 10^6 \text{ M}^{-1} \text{ s}^{-1}$ and $k_d = 0.2 \text{ s}^{-1}$, as reported in the literature for IL-2 binding to IL-2R α (Myszka et al., 1996). A 32×32 grid was used.

0.2 s^{-1} . Analyte was injected for 50 s, after which buffer was introduced and dissociation proceeded for an additional 50 s. Fig. 3 shows the average concentration of bound analyte and the average free analyte concentration just above the sensor surface plotted as a function of time. We see from Fig. 3 that the free analyte concentration near the sensor surface is not constant during the association phase or zero during the dissociation phase, as is often assumed. Fig. 4 presents “snapshots” of the free analyte concentration, as a function of position throughout the flow cell, at four successive time points (two during the association phase and two during the dissociation phase). Note that at 25 s and 75 s there is substantial variation in the free analyte concentration along the sensor surface. As a consequence,




FIGURE 4 Predicted free analyte concentration in the BIACORE flow cell for the simulation parameters given in Fig. 3. Different colors indicate different free analyte concentrations in 0.5 nM increments. The contours separating two colors are lines of fixed concentration. The value of the concentration at any point on the grid is obtained by interpolating between the calculated values of the two nodes the grid line connects. Note that in the association phase ($t = 25$ s and 50 s) the contours separating the two colors nearest the top surface are not smooth. This indicates that in this region, more node points are required to improve accuracy. This simulation was done on a 32×32 grid. Despite the apparent lack of spatial resolution far from the boundary, when we compared calculations using a 32×32 and a 64×64 grid, the quantity we are most interested in, the average bound analyte concentration, \bar{B} , changed only slightly. Between 10 s and 90 s, the difference was always less than 1%.

the bound analyte concentration will also be nonuniform, as illustrated in Fig. 5. In BIACORE instruments the SPR detector is centered in the middle of the flow cell and monitors $\sim 75\%$ of the total flow path. Thus the output is a measure of the average bound analyte over this region.

The parameter values used in Figs. 3–5 were chosen to demonstrate how transport affects the kinetics of binding and dissociation in a BIACORE instrument. In Fig. 6 we show that as the association rate is decreased, transport effects become negligible, i.e., the concentration of free analyte becomes constant in time, and the rapid mixing model becomes a good description of the binding.

In the next section we discuss a proposed method of analyzing BIACORE binding data that are influenced by mass transport. To test this method we simulate additional BIACORE binding experiments under a variety of conditions and see if the method can recover the kinetic parameters used in the simulations.

ANALYZING MASS TRANSPORT-INFLUENCED BINDING DATA

A two-compartment model

Obtaining kinetic parameters from binding data requires fitting a model to the data. At present, it is impractical to do this by using the mathematical description of the previous section, a PDE (Eq. 6), and its boundary conditions, Eqs. 7a–7d. A simpler mathematical model is required, one that can accurately calculate the average concentration of bound analyte at the sensor surface. This model need do nothing more. For example, it does not have to describe the spatial variation of the analyte concentration throughout the flow cell or of the bound analyte over the sensor surface.




FIGURE 5 Spatial variation in the bound analyte concentration, $B(t, x)$. Shown are time courses for the bound analyte concentration in RU at a distance of $l/4$ (—), $l/2$ (---), and $3l/4$ (.....) along the sensor surface. The simulation parameters used are the same as in Fig. 3. The results demonstrate that the bound analyte concentration varies with the distance from the entrance to the flow cell.




FIGURE 6 The time course of the average free analyte concentration just above the sensor surface, for different values of the forward rate constant, k_a . Shown are time courses for the average free analyte concentration for $k_a = 8 \times 10^4 \text{ M}^{-1} \text{ s}^{-1}$ (—), $8 \times 10^5 \text{ M}^{-1} \text{ s}^{-1}$ (---), and $8 \times 10^6 \text{ M}^{-1} \text{ s}^{-1}$ (.....). All other simulation parameters are the same as in Fig. 3.

A two-compartment model, consisting of a set of coupled ordinary differential equations, has been used to analyze a variety of BIACORE binding experiments that have been shown to be influenced by mass transport (Myszka et al., 1996, 1997; Morton and Myszka, 1998). First we review this model, and then in the following section we test it, to see if accurate rate constants can be recovered from simulated BIACORE experiments.

The model handles the variation in analyte concentration in the BIACORE flow cell (see Fig. 4) by dividing the flow chamber into two compartments as shown in Fig. 7. Within each compartment the concentrations are uniform in space but may change in time. These concentrations are averages over the length of the BIACORE flow cell. It is assumed that the analyte concentration in the outer compartment is




FIGURE 7 A two-compartment model of the BIACORE flow cell. In the binding phase the average analyte concentration near the sensor surface, C , equals zero initially and grows with time until it equals the injection concentration C_T or dissociation is initiated and it decays to zero. The analyte concentration far from the sensor surface remains constant and equal to C_T during binding, and constant and equal to zero during dissociation. B and R are the bound and free receptor concentrations, averaged over the sensor surface.

constant and equal to the injection concentration, C_T . The model ignores the brief transition periods that occur when binding or dissociation is initiated and the concentration in the outer compartment rises or falls to C_T . At the typical flow rates used in kinetic experiments (50–100 $\mu\text{l}/\text{min}$, corresponding to $v_c = 5 - 10 \text{ cm/s}$), the time needed for an analyte molecule that is not near the walls of the flow cell to travel 0.24 cm, the length of the flow chamber, is less than 0.05 s. The concentration in the inner compartment, C , changes because analyte is transported between compartments and because analyte binds to, and dissociates from, immobilized receptors on the sensor surface. If V_i is the volume of the inner compartment, S is the surface area of the flow chamber, and k_M is the transport coefficient describing diffusive movement of analyte between the compartments, then

$$V_i dC/dt = S(-k_a C(R_T - B) + k_d B + k_M(C_T - C)) \quad (8a)$$

$$dB/dt = k_a C(R_T - B) - k_d B \quad (8b)$$

The initial conditions for the binding phase of the experiment are that at $t = 0$, C_T equals the injection concentration, $C = 0$ and $B = 0$. The initial conditions for the dissociation phase are that at the time of dissociation, t_{diss} , $C_T = 0$, and C and B equal the values they attained at $t = t_{\text{diss}}$. If the association phase is sufficiently long so that a steady state is reached, then at $t = t_{\text{diss}}$, $C = C_T$ and $B = KC_T R_T / (1 + KC_T)$, but this is not required.

The rapid mixing model assumes that after a brief transient, the free analyte concentration is uniform throughout the flow chamber. This model is described by Eq. 8b with $C = C_T$.

As discussed in the Appendix, k_M is the diffusion-limited forward rate constant, averaged over the sensor surface. Transport effects will influence the kinetics of binding when the reaction rate is fast compared to the rate of transport, i.e., when $k_a R_T \geq k_M$. (To evaluate this inequality, the units must be consistent, so for example, if k_a is in $\text{M}^{-1} \text{s}^{-1}$, then R_T should be in M cm and k_M in cm/s .) As discussed by Lok et al. (1983) and Sjölander and Urbaniczky (1991), it can be shown that, to a good approximation,

$$k_M = \frac{1}{\Gamma(4/3)} \left(\frac{3v_c D^2}{2hl} \right)^{1/3} \approx 1.282 \left(\frac{v_c D^2}{hl} \right)^{1/3} \quad (9)$$

Because the model does not specify the height of the inner compartment, there are four unknown parameters that enter Eqs. 8a and 8b, k_a , k_d , k_M , and $h_i = V_i/S$. However, for the range of parameter values that describe BIACORE experiments, the solutions to Eqs. 8a and 8b are insensitive to the value of h_i . It can be shown (see Appendix and Fig. 3) that at the sensor surface, once binding or dissociation is initiated, C changes rapidly over a short period of time and slowly thereafter. Furthermore, while this rapid change in C is occurring, there is a negligible change in B . This behavior means that a quasi-steady-state approximation can be made where we set $dC/dt = 0$ (Segel and Slemrod, 1989). When

this is done, we see from Eq. 8a that h_i drops out of the equations, which explains why the choice of h_i does not affect the solution. Once the quasi-equilibrium approximation is made, the model is identical to the often discussed effective rate constant model (Karlsson et al., 1994).

Returning to Eqs. 8a and 8b, we define $\tilde{B} = B/h_i$, $\tilde{R} = R/h_i$, $\tilde{R}_T = R_T/h_i$, and $\tilde{k}_M = k_M/h_i$. In terms of these quantities, Eqs. 8a and 8b take the familiar form

$$dC/dt = (-k_a C(\tilde{R}_T - \tilde{B}) + k_d \tilde{B} + \tilde{k}_M(C_T - C)) \quad (10a)$$

$$d\tilde{B}/dt = k_a C(\tilde{R}_T - \tilde{B}) - k_d \tilde{B} \quad (10b)$$

When these equations are used to analyze binding data, some care must be taken in the choice of units. If we measure R , R_T , and B in RU; C and C_T in M; k_a in $\text{M}^{-1} \text{s}^{-1}$; k_d in s^{-1} ; and k_M in cm/s , this is equivalent to setting $h_i = 1 \text{ RU/M}$. (Because the data are insensitive to the value of h_i , we are free to choose any value for it we wish. Taking $h_i = 1 \text{ RU/M}$ avoids having to divide the data values by h_i before fitting the data. It is a convenient choice having nothing to do with the physical height of the inner compartment.) To obtain k_M in cm/s from \tilde{k}_M in $1/\text{s}$, the parameter determined from the least-squares fit of the data, one must multiply by 10^{-7} and divide by the molecular weight of the analyte, i.e.,

$$\begin{aligned} k_M &= (1 \text{ RU/M})\tilde{k}_M = (10^{-7} \text{ cm g/mol})\tilde{k}_M \\ &= (10^{-7} \text{ cm/MW})\tilde{k}_M \end{aligned} \quad (11)$$

where we have used the conversion factor $1 \text{ RU} = 10^{-10} \text{ g/cm}^2$.

Data fitting approach

We briefly summarize the data fitting approach we employ, because it has been discussed in detail elsewhere (Myszka, 1997; Myszka et al., 1997; Morton and Myszka, 1998). The rate constants k_a and k_d , \tilde{k}_M (related to the transport coefficient k_M by Eq. 11), and the surface receptor density R_T are taken to be free parameters. The best fit values of these parameters are determined from nonlinear least-squares fitting of the data, where the calculated values of the average bound analyte concentration for each time point are determined by solving Eqs. 10a and 10b numerically (Morton et al., 1995). (The software program used in this analysis (CLAMP) may be downloaded for no charge at the following web site: <http://www.hci.utah.edu/cores/biacore/docs/clamp.html>.)

To obtain accurate values for the kinetic rate constants and the transport coefficient, the data that are being fit must be sensitive to these parameters. Recall that as the receptor density on the sensor surface is increased, the binding reaction at the sensor surface speeds up, and the binding kinetics become transport limited. As a result, at high surface receptor densities, the binding kinetics are dependent on k_M . In the other limit, as the receptor density is reduced,

transport effects decrease and these data become insensitive to k_M and strongly dependent on the reaction rate constants. To try to ensure that the data depend on both the transport and the reaction rate constants, we simultaneously fit data sets for different surface receptor densities. The procedure is sometimes called global analysis and has been extensively used to analyze BIACORE binding data (Morton et al., 1995; Myszka et al., 1996, 1997; Morton and Myszka, 1998).

RESULTS

In this section we report the results of fitting the two-compartment model to simulated data sets generated under a variety of conditions. Before considering the effects of noisy data on parameter estimation, we see how well the model does at recovering parameter values in the absence of noise.

We start by analyzing data where the reactions are not influenced by mass transport. Fig. 8 illustrates two sets of binding experiments simulated with identical rate constants ($k_a = 5 \times 10^4 \text{ M}^{-1} \text{ s}^{-1}$ and $k_d = 0.04 \text{ s}^{-1}$) and flow rates (100 $\mu\text{l}/\text{min}$), but with different receptor densities ($R_{T1} = 50 \text{ RU}$ and $R_{T2} = 15 \text{ RU}$). Note that when the reaction is not influenced by transport, for the same analyte concentration, the binding responses are identical for the two surfaces when normalized with respect to the maximum response, i.e., a plot of the fraction of receptor sites bound versus time is independent of R_T . Under these conditions, the rapid mixing model (Eq. 8b, with $C = C_T$) gives an excellent fit to the data (not shown) and returns correct parameter values, as shown in Table 1A, i.e., the best fit values of the parameters match the values used to simulate the data. The standard deviation of the residuals equals 0.054 RU, which is small but nonzero. Because the ODE models used to fit the data are always approximations to the PDE model used to

simulate the data, even in the absence of noise, the standard deviation of the residuals is nonzero. As shown in Table 1B, applying the two-compartment model to these same data sets gives essentially the same results. The additional parameter, \tilde{k}_M , has little effect on the quality of the fit (std residuals = 0.053 RU), the returned parameter values, standard deviations, or the correlation coefficients. However, the value returned for \tilde{k}_M of $2.9 \times 10^{12} \text{ s}^{-1}$, which from Eq. 11 corresponds to $k_M = 9.7 \text{ cm/s}$ for $\text{MW} = 3 \times 10^4 \text{ Da}$, is grossly in error. From Eq. 9, the predicted value for k_M is $2.6 \times 10^{-3} \text{ cm/s}$. This error is not surprising. The association rate is so slow that the binding is unaffected by transport, and so the data contain no information for the determination of k_M .

Next we carried out the same fitting procedures on data simulated with a 10-fold higher association rate constant ($k_a = 5 \times 10^5 \text{ M}^{-1} \text{ s}^{-1}$). All other parameters were the same. Again a good fit was obtained with the rapid mixing model (std residuals = 0.062 RU), yielding parameter values that are close to the input values (Table 2A). As shown in Table 2B, using a two-compartment model improves the quality of the fit somewhat (std residuals = 0.055 RU). Furthermore, the best fit value $\tilde{k}_M = 8.9 \times 10^8 \text{ s}^{-1}$, which corresponds to $k_M = 3.0 \times 10^{-3} \text{ cm/s}$, is greatly improved, being within 13% of the expected value. The improvements in the residual standard deviation, rate constant estimates, and the value of k_M suggest that under these conditions, transport is beginning to influence the binding kinetics.

Fig. 9 illustrates what happens to the shapes of the binding responses when the association rate constant is increased to $1 \times 10^8 \text{ M}^{-1} \text{ s}^{-1}$ and the data become strongly influenced by transport. (Note that the off-rate constant was increased to 0.08 s^{-1} , so that the shapes of the progress curves would remain similar.) Without doing any data fitting, one can see (compare Fig. 9 c with 9 d) that the binding is transport limited because, unlike in Fig. 8, the shapes of




FIGURE 8 Binding response simulated for a reaction occurring under conditions where the binding is reaction limited. The parameter values used in the simulations were $k_a = 5 \times 10^4 \text{ M}^{-1} \text{ s}^{-1}$, $k_d = 0.04 \text{ s}^{-1}$, $D = 1 \times 10^{-6} \text{ cm}^2/\text{s}$, and $v_c = 10 \text{ cm/s}$. The surface receptor concentrations were 0.167 nM cm (a) and 0.05 nM cm (b), which, for an analyte with $\text{MW} = 3 \times 10^4$, correspond to a maximum response of 50 RU and 15 RU, respectively. The ligand concentrations used in the simulation were 8000, 2670, 890, 297, and 99 nM. The results of the fit are given in Table 1.

TABLE 1 Parameter estimates, standard deviations, and correlations for the fits of the rapid mixing model and the two-compartment model to data simulated in Fig. 8 with $k_a = 5 \times 10^4 \text{ M}^{-1} \text{ s}^{-1}$ and $k_d = 4 \times 10^{-2} \text{ s}^{-1}$

Parameter	Estimate	Std	Correlation coefficient			
			$B_{\max 1}$	$B_{\max 2}$	k_a	k_d
A. Rapid mixing model						
R_{T1} (RU)	50.14	8.6×10^{-3}				
R_{T2} (RU)	15.00	5.5×10^{-3}	0.324			
k_a ($\text{M}^{-1} \text{ s}^{-1}$)	4.92×10^4	2.7×10^1	-0.632	-0.319		
k_d (s^{-1})	0.040	1.3×10^{-3}	0.342	0.165	0.230	
B. Two-compartment model						
R_{T1} (RU)	50.14	8.6×10^{-3}				
R_{T2} (RU)	15.00	5.5×10^{-3}	0.323			
k_a ($\text{M}^{-1} \text{ s}^{-1}$)	4.92×10^4	2.7×10^1	-0.631	-0.317		
k_d (s^{-1})	0.040	1.3×10^{-3}	0.343	0.161	0.226	
k_M (s^{-1})	2.89×10^{12}	1.2×10^{12}	-0.009	0.044	0.020	-0.090

The parameters values used to simulate the data are given in Fig. 8. $k_a R_T / k_M = 0.0032$ and 0.0010 , for the two surface receptor densities $1.67 \times 10^{-10} \text{ M cm}$ and $5.0 \times 10^{-9} \text{ M cm}$. For an analyte with a molecular weight of 3×10^4 , these correspond to maximum binding capacities of 50 RU and 15 RU.

the binding curves for the two surfaces are not identical. Now equilibrium is approached more rapidly on the surface with the lower receptor density. When the binding is influenced by transport, the fraction of receptor sites bound is no longer independent of R_T . Receptors no longer act independently. In the association phase they compete for analyte, whereas in the dissociation phase rebinding occurs.

The fit of the rapid mixing model to the simulated data (Fig. 9, *a* and *b*) is poor (residual std = 1.52 RU), whereas the fit of the two-compartment model (Fig. 9, *c* and *d*) is excellent (residual std = 0.07 RU). Furthermore, in Table 3A we see that the two-compartment model returns parameter values that are close to those used in the simulations. The best fit value $\tilde{k}_M = 7.72 \times 10^8 \text{ RU s}^{-1}$, or equivalently, $k_M = 2.6 \times 10^{-3} \text{ cm/s}$, is within 1% of the predicted value. These results demonstrate that in the absence of noise, the two-compartment model returns accurate estimates for both the intrinsic reaction rate constants and the transport coefficient when the binding kinetics are influenced by transport.

efficient when the binding kinetics are influenced by transport.

The accurate determination of k_M suggests that BIA-CORE can be used to determine the diffusion coefficient of the analyte by calculating D from Eq. 9. The value of \tilde{k}_M in Table 3A gives $D = 9.85 \times 10^{-7} \text{ cm}^2/\text{s}$, which is remarkably close to the value used in the simulation of $1.0 \times 10^{-6} \text{ cm}^2/\text{s}$. However, because Eq. 9 is an approximate result, to determine D from k_M , one must ensure that the experiments are done in a parameter range in which Eq. 9 is sufficiently accurate (see Discussion).


Finally, to study the effects of noise on the best fit values of the parameters, we added 0.5 RU of pseudo-normally distributed noise to the data sets of Fig. 9. This level of noise is consistent with what is seen in typical experiments done on BIACORE 2000 (Morton and Myszk, 1998). Fig. 10 shows the results of fitting the two-compartment model to these data. The residual standard deviation was 0.504 RU,

TABLE 2 Parameter estimates, standard deviations, and correlations for the fits of the rapid mixing model and the two-compartment model to data simulated with $k_a = 5 \times 10^5 \text{ M}^{-1} \text{ s}^{-1}$ and $k_d = 4 \times 10^{-2} \text{ s}^{-1}$

Parameter	Estimate	Std	Correlation coefficient			
			R_{T1}	R_{T2}	k_a	k_d
A. Rapid mixing model						
R_{T1} (RU)	50.12	9.6×10^{-3}				
R_{T2} (RU)	15.00	6.3×10^{-3}	0.291			
k_a ($\text{M}^{-1} \text{ s}^{-1}$)	4.86×10^5	2.8×10^2	-0.601	-0.277		
k_d (s^{-1})	0.039	1.5×10^{-3}	0.335	0.154	0.277	
B. Two-compartment model						
R_{T1} (RU)	50.13	8.6×10^{-3}				
R_{T2} (RU)	15.00	5.6×10^{-3}	0.293			
k_a ($\text{M}^{-1} \text{ s}^{-1}$)	4.94×10^5	5.4×10^2	-0.338	-0.164		
k_d (s^{-1})	0.040	3.6×10^{-3}	0.061	0.019	0.870	
k_M (s^{-1})	8.93×10^8	5.1×10^7	0.064	0.039	-0.881	-0.930

The parameters values used to simulate the data are given in Fig. 8, except for $k_a = 5 \times 10^5 \text{ M}^{-1} \text{ s}^{-1}$ and $k_d = 4 \times 10^{-2} \text{ s}^{-1}$. $k_a R_T / k_M = 0.032$ and 0.010 , for the two surface receptor densities $1.67 \times 10^{-10} \text{ M cm}$ and $5.0 \times 10^{-9} \text{ M cm}$. For an analyte with a molecular weight of 3×10^4 , these correspond to maximum binding capacities of 50 RU and 15 RU.

FIGURE 9 Binding response simulated for a reaction occurring under conditions where the binding is transport limited. The parameter values used in the simulations were $k_a = 1 \times 10^8 \text{ M}^{-1} \text{ s}^{-1}$ and $k_d = 0.08 \text{ s}^{-1}$. All other parameters were the same as in Fig. 8. The surface receptor concentrations correspond to a maximum response of 50 RU (a, c) and 15 RU (b, d) for a 3×10^4 Da analyte. The analyte concentrations used in the simulation were 8, 2.67, 0.89, 0.297, and 0.099 nM. Shown are simultaneous fits (red lines) of the rapid mixing model (a, b) and the two-compartment model (c, d) to the simulations. The results of the fits are given in Table 3.



which is consistent with the level of noise added to the data. The best fit parameter values are given in Table 3B. Surprisingly, this level of random noise had little effect on the determined parameter values or the correlation coefficients, although as expected, the standard errors were increased. To


demonstrate that the linear approximation statistics returned for the nonlinear analysis were valid, we performed Monte Carlo analysis to generate confidence probability distributions for the estimated parameters (Straume and Johnson, 1992). One thousand data sets were simulated with noise

TABLE 3 Parameter estimates, standard deviations, and correlations for the fits of the two-compartment model to data simulated with $k_a = 1 \times 10^8 \text{ M}^{-1} \text{ s}^{-1}$ and $k_d = 8 \times 10^{-2} \text{ s}^{-1}$ in the absence of noise (Fig. 9) and the presence of noise (Fig. 10)

Parameter	Estimate	Std	Correlation coefficient			
			R_{T1}	R_{T2}	k_a	k_d
A. Absence of noise						
$B_{\max1}$ (RU)	49.92	1.2×10^{-2}				
$B_{\max2}$ (RU)	15.04	7.5×10^{-3}	0.349			
k_a ($\text{M}^{-1} \text{ s}^{-1}$)	1.02×10^8	4.1×10^5	-0.156	0.014		
k_d (s^{-1})	0.083	3.1×10^{-4}	0.012	0.102	0.979	
k_M (s^{-1})	7.72×10^8	9.2×10^5	0.078	-0.049	-0.894	-0.898
B. Noisy data						
R_{T1} (RU)	49.90	8.7×10^{-2}				
R_{T2} (RU)	14.99	5.3×10^{-2}	0.348			
k_a ($\text{M}^{-1} \text{ s}^{-1}$)	1.05×10^8	3.0×10^6	-0.168	0.021		
k_d (s^{-1})	0.085	2.3×10^{-3}	0.002	0.107	0.980	
k_M (s^{-1})	7.66×10^8	6.4×10^6	0.086	-0.055	-0.894	-0.897
C. Monte Carlo analysis after 1000 simulations						
R_{T1} (RU)	49.92	8.7×10^{-2}				
R_{T2} (RU)	15.03	5.3×10^{-2}				
k_a ($\text{M}^{-1} \text{ s}^{-1}$)	1.02×10^8	2.9×10^6				
k_d (s^{-1})	0.083	2.2×10^{-3}				
k_M (s^{-1})	7.72×10^8	6.7×10^6				

The parameters values used to simulate the data are given in Figs. 9 and 10. $k_a R_T / k_M = 6.4$ and 1.9, for the two surface receptor densities 1.67×10^{-10} M cm and 5.0×10^{-9} M cm. For an analyte with a molecular weight of 3×10^4 , these correspond to maximum binding capacities of 50 RU and 15 RU.

FIGURE 10 Simulations of the binding response of Fig. 9 in the presence of noise. Random noise with a standard deviation of 0.5 RU was generated and added to the simulations of Fig. 9. As in Fig. 9, the surface receptor concentrations correspond to a maximum response of 50 RU (a) and 15 RU (b) for a 3×10^4 Da analyte. Shown are simultaneous fits (red lines) of the two-compartment model to the simulations. The results of the fits are given in Table 3.



and solved as before with varied starting parameters. The results of this analysis confirmed that the confidence intervals were normally distributed (data not shown), and the standard deviations for the confidence intervals (Table 3C) were similar to those determined from the linear approximation statistics (Table 3B). The Monte Carlo analysis routines are readily available in the CLAMP program.

DISCUSSION

The binding kinetics in a BIACORE flow cell are the result of diffusion and flow, which bring the analyte to, and take it away from, the sensor surface, and the chemical reaction that occurs between analyte and receptor at the sensor surface. We have modeled these processes for the case of a monovalent analyte interacting with a monovalent receptor, when the conditions under which the receptor is immobilized do not affect the binding kinetics. If, for example, the receptor is coupled to the sensor surface through a dextran layer, then our model is restricted to conditions where the layer has no influence on the binding kinetics. As discussed in the Appendix, numerous authors have considered this case and argued that a two-compartment model gives a reasonable description of the binding kinetics when the data are influenced by transport (Glaser, 1993; Karlsson et al., 1994; Schuck and Minton, 1996; Christensen, 1997), although its use in analyzing such data has been questioned (Schuck, 1997b). To see if, in fact, a two-compartment model can be used to extract accurate parameter values from transport-influenced BIACORE binding data, we generated test sets of simulated data. The obvious advantage of using simulated data sets is that the values of the underlying parameters are known exactly. We extended the model of Glaser (1993) to include diffusion in the direction of flow, obtaining a partial differential equation with appropriate boundary conditions that accounts for the processes that occur in a flow cell: convection, diffusion, and reversible chemical reaction at the sensor surface. The simulated data sets were generated by solving this model numerically, to obtain noiseless data, and then adding noise to the numerical solutions. These simulated data sets served as our standard for testing the two-compartment model. We showed that

when the two-compartment model was fit to the simulated data, values for the rate constants were returned with surprisingly high accuracy.

The data we simulated corresponded to optimal experimental conditions currently available on BIACORE 2000, a flow rate of 100 $\mu\text{l}/\text{min}$ and surface receptor densities as low as 15 RU. Our data analysis method involved simultaneously fitting the two-compartment model to two sets of simulated data, corresponding to two different receptor densities, 15 RU and 50 RU. The idea behind the approach, known as global analysis (reviewed in Myszka et al., 1997), is to obtain data that are sensitive to both the chemical rate constants (data from the low-capacity surface) and the transport coefficient (data from the high-capacity surface). In our fits we showed that even though the maximum response from one of the surfaces was only 15 RU, the addition of the expected level of experimental noise (0.5 RU) did not affect our ability to accurately recover the reaction rate constants.

It has been pointed out that if receptors are immobilized in a dextran layer, the binding kinetics can be influenced by the transport of the analyte within the layer (Schuck, 1996; Yarmush et al., 1996). It is expected that such effects will become pronounced at high receptor densities. At such densities the two-compartment model will no longer give a reasonable description of the binding kinetics. Elsewhere we discuss how to reformulate the model when the thickness of the layer cannot be neglected (Wofsy et al., manuscript in preparation). There we show that the effects of the layer will be negligible when the layer "looks" thin to the analyte, i.e., when the average distance the analyte travels in the layer before it binds to a receptor (its mean free path) is long compared to the height of the layer, d . The mean free path equals $\sqrt{(dD_i)/(k_a R_T)}$, where R_T/d is the three-dimensional receptor concentration and D_i is the diffusion coefficient in the layer. Thus one can ignore the layer and treat the binding as if it is directly to the sensor surface when $1 \gg \sqrt{k_a d R_T / D_i}$. (For the low receptor densities, we expect $D_i \approx D$. In our simulations the 50 RU surface corresponded to $R_T = 0.167$ nM cm and $D = 1 \times 10^{-6}$ cm²/s. For a standard flow cell, $d \approx 1 \times 10^{-5}$ cm. For these values the inequality breaks down when $k_a \geq 1 \times 10^8$ M⁻¹ s⁻¹.)

In addition to determining the rate constants, we showed that the two-compartment model can be used to determine D , the diffusion coefficient of the analyte. This was done by determining the transport coefficient, k_M , by fitting the simulated data, and then calculating D from Eq. 9. Because Eq. 9 is an approximate result, it can be used to determine D only under conditions in which it is highly accurate. Equation 9 overestimates k_M . It is derived by ignoring the parabolic flow velocity profile and taking the velocity to rise linearly with the height above the sensor surface. It also ignores the upper boundary of the flow cell. Analyte that diffuses to this boundary is not reflected. However, if the time it takes to travel the length of the flow cell, $l/\bar{v} = 3l/(2v_c)$, is short compared to the time it takes to diffuse to the boundary, $h^2/(4D)$, or equivalently, if $1 \gg 6lD/(h^2v_c)$, then we expect Eq. 6 to be valid. Thus to determine D from Eq. 9, one should use a fast flow velocity for the binding studies.

We have shown that the two-compartment model can be used to determine accurate values of the rate constants and transport coefficients for data that are influenced by flow and diffusion. From the transport coefficient the analyte diffusion coefficient can be determined as well. Using the two-compartment model and analyzing the data globally extends the range of reaction rates that can be determined with instruments like BIACORE. At present, this approach seems to be the best available for analyzing experiments when the effects of mass transport cannot be eliminated from the binding kinetics. This model has been incorporated in BIAevaluation 3.0 analysis software available from BIA-CORE, Inc. It has already been used to analyze such experimental data, yielding excellent fits to the data and improved accuracy of the returned rate constants (Myszka et al., 1996; Myszka et al., 1997; Morton and Myszka, 1998).

APPENDIX

The two-compartment model is given by Eqs. 8a and 8b. If we make a quasi-steady-state approximation and set $dC/dt = 0$, Eq. 8a becomes

$$C = \frac{k_d B + k_M C_T}{k_a(R_T - B) + k_M} \quad (\text{A1})$$

Substituting this expression for C into Eq. 8b, we obtain

$$\begin{aligned} dB/dt = & \frac{k_a}{1 + k_a(R_T - B)/k_M} C_T(R_T - B) \\ & - \frac{k_d}{1 + k_a(R_T - B)/k_M} B \end{aligned} \quad (\text{A2})$$

Equation A2 has the same form as a well-mixed system, except that the true rate constants are replaced by effective rate coefficients that change in time as the free receptor concentration, $R = R_T - B$, changes. If this equation is to have the correct limit as $R \rightarrow \infty$, i.e., when the sensor surface acts as a perfect absorber, then k_M must equal the diffusion-limited forward rate constant averaged over the sensor surface (Eq. 9). Numerous authors have presented Eq. A2 and discussed its implications (Glaser, 1993; Karlsson et al., 1994; Schuck and Minton, 1996; Christensen, 1997).

Justifying the quasi-steady-state approximation that leads to Eq. A2 requires showing that when binding or dissociation is initiated, C at the sensor surface changes rapidly over a short period of time and slowly thereafter. Furthermore, if we take the initial condition for Eq. A2 to be that $B = 0$ at $t = 0$, then we must show that B changes negligibly during the rapid transient (Segel and Slemrod, 1989). Because this is the behavior we see in our simulations (e.g., the *inset* in Fig. 3), we expect that the approximation is justified.

To find the characteristic time for the fast transient in C , t_f , during the association phase, we assume that $B \approx 0$ during this time and justify this assumption after finding t_f . (For the fast transient during the dissociation phase, we would assume that B remains constant.) At the start of binding, when $B \approx 0$, Eq. 8a is approximately

$$h_i dC/dt \approx k_M C_T - (k_M + k_a R_T) C \quad (\text{A4})$$

with solution

$$C = C_T (k_M / (k_M + k_a R_T)) (1 - e^{-t/t_f}) \quad (\text{A5})$$

where

$$t_f = h_i / (k_M + k_a R_T) < h / (k_M + k_a R_T) < h / k_M \quad (\text{A6})$$

For the simulations in Figs. 8–10, $k_M = 2.6 \times 10^{-3}$ cm/s. For a standard flow cell, $h = 5 \times 10^{-3}$ cm, so we have for these simulations that $t_f < 2$ s, i.e., C changes rapidly for less than 2 s and then slowly until dissociation is initiated. During this time the maximum amount of analyte that can bind is $B_{\max} = k_a C_T R_T t_f$, and this expression can be evaluated to see if the binding is significant. We simply note that in kinetic binding studies using BIACORE, analyte concentrations are chosen so that binding occurs over much longer periods than a few seconds; thus we expect the binding to be negligible in the first second or so.

The authors acknowledge helpful discussions with C. Wofsy.

This work was supported by National Institutes of Health Grant GM35556 and performed under the auspices of the U.S. Department of Energy.

REFERENCES

- Batchelor, G. K. 1967. *An Introduction to Fluid Dynamics*. Cambridge University Press, Cambridge.
- Brody, J. P., P. Yager, R. E. Goldstein, and R. H. Austin. 1996. Biotechnology at low Reynolds number. *Biophys. J.* 71:3430–3441.
- Chaiken, I., S. Rosé, and R. Karlsson. 1992. Analysis of macromolecular interactions using immobilized ligands. *Anal. Biochem.* 201:197–210.
- Christensen, L. L. H. 1997. Theoretical analysis of protein concentration determination using biosensor technology under conditions of partial mass transport limitation. *Anal. Biochem.* 249:153–164.
- Dembo, M. 1994a. On free boundary problems and amoeboid motion. *In Biomechanics of Active Movement and Division of Cells*. N. Akkas, editor. Springer-Verlag, Berlin. 231–238.
- Dembo, M. 1994b. Continuum theories of cytoskeletal mechanics: solution by a finite element method. Los Alamos Unclassified Report 94-3454.
- Fivash, M., E. M. Towler, and R. J. Fisher. 1998. BIAcore for macromolecular interactions. *Curr. Opin. Biotechnol.* 9:97–101.
- Garland, P. B. 1996. Optical evanescent wave methods for the study of biomolecular interactions. *Q. Rev. Biophys.* 29:91–117.
- Glaser, R. W. 1993. Antigen-antibody binding and mass transport by convection and diffusion to a surface: a two-dimensional computer model of binding and dissociation kinetics. *Anal. Biochem.* 213:152–161.
- Ito, W., and Y. Kurosawa. 1993. Development of an artificial antibody system with multiple valency using an Fv fragment fused to a fragment protein A. *J. Biol. Chem.* 268:20668–20675.
- Jönsson, U., L. Fagerstam, B. Ivarsson, B. Johnsson, R. Karlsson, K. Lundh, S. Lofas, B. Persson, H. Roos, I. Ronnberg, R. Sjolander, E.

- Stenberg, R. Stahlberg, C., Urbaniczky, H. Ostlin, and M. Malmqvist. 1991. Real-time biospecific interaction analysis using surface plasmon resonance and a sensor chip technology. *BioTechniques*. 11:620–627.
- Karlsson, R., and A. Fält. 1997. Experimental design for improved kinetic analysis of protein-protein interactions with surface plasmon resonance biosensors. *J. Immunol. Methods*. 200:121–133.
- Karlsson, R., H. Roos, L. Fägerstam, and B. Persson. 1994. Kinetic and concentration analysis using BIA technology. *Methods: A Companion to Methods Enzymol.* 6:99–110.
- Lok, B. K., Y.-L. Cheng, and C. R. Robertson. 1983. Protein adsorption on crosslinked polydimethylsiloxane using total internal reflection fluorescence. *J. Colloid Interface Sci.* 91:104–116.
- Malmborg, A.-C., A. Michaëlsson, M. Ohlin, B. Jansson, and C. A. K. Borrebaeck. 1992. Real time analysis of antibody-antigen reaction kinetics. *Scand. J. Immunol.* 35:643–650.
- Malmqvist, M. 1993. Surface plasmon resonance for detection and measurement of antibody-antigen affinity and kinetics. *Curr. Opin. Immunol.* 5:282–286.
- Morton, T. A., D. B. Bennett, E. R. Appelbaum, D. M. Cusimano, K. O. Johanson, R. E. Matico, P. R. Young, M. Doyle, and I. M. Chaiken. 1994. Analysis of the interaction between human interleukin-5 and the soluble domain of its receptor using a surface plasmon resonance biosensor. *J. Mol. Recognit.* 7:47–55.
- Morton, T. A., and D. G. Myszka. 1998. Improved methods for kinetic analysis of macromolecular interactions using surface plasmon resonance biosensors. *Methods Enzymol.* (in press).
- Morton, T. A., D. G. Myszka, and I. M. Chaiken. 1995. Interpreting complex binding kinetics from optical biosensors: a comparison of analysis by linearization, the integrated rate equation, and numerical integration. *Anal. Biochem.* 227:176–185.
- Myszka, D. G. 1997. Kinetic analysis of macromolecular interactions using surface plasmon resonance biosensors. *Curr. Opin. Biotechnol.* 8:50–57.
- Myszka, D. G., P. R. Arulanantham, Z. Wu, T. Sana, and T. L. Ciardelli. 1996. Kinetic analysis of ligand binding to interleukin-2 receptor complexes created on an optical biosensor surface. *Protein Sci.* 5:2468–2478.
- Myszka, D. G., T. A. Morton, M. L. Doyle, and I. M. Chaiken. 1997. Kinetic analysis of a protein antigen-antibody interaction limited by mass transport on an optical biosensor. *Biophys. Chem.* 64:127–137.
- Parsons, I. D., and P. G. Stockley. 1997. Quantitation of the *Escherichia coli* methionine repressor-operator interaction by surface plasmon resonance is not affected by the presence of a dextran matrix. *Anal. Biochem.* 254:82–87.
- Raghavan, M., and P. J. Bjorkman. 1995. BIAcore: a microchip-based system for analyzing the formation of macromolecular complexes. *Structure*. 3:331–333.
- Roden, L. D., and D. G. Myszka. 1996. Global analysis of a macromolecular interaction measured on BIACORE. *Biochem. Biophys. Res. Commun.* 225:1073–1077.
- Schuck, P. 1996. Kinetics of ligand binding to receptor immobilized in a polymer matrix, as detected with an evanescent wave biosensor. I. A computer simulation of the influence of mass transport. *Biophys. J.* 70:1230–1249.
- Schuck, P. 1997a. Use of surface plasmon resonance to probe the equilibrium and dynamic aspects of interactions between biological macromolecules. *Annu. Rev. Biophys. Biomol. Struct.* 26:541–566.
- Schuck, P. 1997b. Reliable determination of binding affinity and kinetics using surface plasmon resonance biosensors. *Curr. Opin. Biotechnol.* 8:498–502.
- Schuck, P., and A. P. Minton. 1996. Analysis of mass transport-limited binding kinetics in evanescent wave biosensors. *Anal. Biochem.* 240:262–272.
- Segel, L. A., and M. Slemrod. 1989. The quasi-steady-state assumption: a case study in perturbation. *SIAM Rev.* 31:446–477.
- Silin, V., and A. Plant. 1997. Biotechnological applications of surface plasmon resonance. *Trends Biotechnol.* 15:353–359.
- Sjölander, S., and C. Urbaniczky. 1991. Integrated fluid handling system for biomolecular interaction analysis. *Anal. Chem.* 63:2338–2345.
- Stenberg, E., B. Persson, H. Roos, and C. Urbaniczky. 1991. Quantitative determination of surface concentration of protein with surface plasmon resonance using radiolabeled proteins. *J. Colloid Interface Sci.* 143:513–526.
- Straume, M., and M. L. Johnson. 1992. Monte Carlo method for determining complete confidence probability distributions of estimated model parameters. *Methods Enzymol.* 210:117–129.
- Szabo, A., L. Stolz, and R. Granzow. 1995. Surface plasmon resonance and its use in biomolecular interaction analysis (BIA). *Curr. Opin. Struct. Biol.* 5:699–705.
- van der Merwe, P. A., and A. N. Barclay. 1996. Analysis of cell-adhesion molecule interactions using surface plasmon resonance. *Curr. Opin. Immunol.* 8:257–261.
- Wohlueter, R. M., K. Parekh, V. Udhayakumar, S. Fang, and A. A. Lal. 1994. Analysis of binding of monoclonal antibody to a malarial peptide by surface plasmon resonance biosensor and integrated rate equations. *J. Immunol.* 153:181–189.
- Yarmush, M. L., D. B. Patankar, and D. M. Yarmush. 1996. An analysis of transport resistances in the operation of BIACORE: implications for kinetic studies of biospecific interactions. *Mol. Immunol.* 33:1203–1214.